

BLINK FRAMES

Indications for Use

This product is designated as a prescription device. It is intended to hold prescription and non-prescription lenses, and be worn by the individual it was prescribed to. May be indicated for the capture and storage of images of the eye and surrounding area in the general population.

 **Refer to instructions for use.**

Document Location

This user manual may be found on the About page in the settings menu of the Blink Frames Application or at <https://blinkframes.com/usermanual>

Please complete and retain for your records

The frames' serial number, as well as the model/reference number, are both located on the shipping box.

Serial number:

Model number:



Date of Manufacture: The first four digits in the serial number indicates the year of manufacture.



AT A GLANCE

Indications for Use.....	2
--------------------------	---

IMPORTANT SAFETY INSTRUCTIONS

Important Safety Instructions	6
-------------------------------------	---

For Healthcare Professionals	6
------------------------------------	---

For Patients and Healthcare Professionals	7
---	---

RESTRICTIONS FOR USE	8
-----------------------------------	---

ESSENTIAL PERFORMANCE	9
------------------------------------	---

REGULATORY AND LEGAL INFORMATION	10
---	----

LICENSE DISCLOSURES	12
----------------------------------	----

INTRODUCTION

Blink Frames System	15
---------------------------	----

Information on Patch Wearing	15
------------------------------------	----

Support Materials	15
-------------------------	----

GENERAL INFORMATION

Sweat and Water Resistance	16
----------------------------------	----

Limited Warranty	16
------------------------	----

Lifetime	16
----------------	----

Model/Reference Number Locations	17
--	----

WHAT'S IN THE BOX

Contents	18
----------------	----

BLINK FRAMES APP SETUP

Download the Blink Frames App	19
-------------------------------------	----

Add the Frames to an Existing Account	19
---	----

BATTERY

Charge the Frames	20
Check the Frame's Battery Level	20
View the Battery Level	20

BLUETOOTH CONNECTIONS

Connecting to Your Frames on Your Mobile Device	21
Disconnect a Mobile Device	22
Reconnect a Mobile Device	22
Auto-Connect	22
WiFi	22
De-authenticate Frames	22
Wipe Frame Memory	22
Reboot Frames	22

FRAME STATUS

Bluetooth Connection Status	23
Battery Status	23
Device Status	23
In-App Notifications	24
Notification Icons	24
Frame Notifications	24

FRAME CONTROLS

Tap and Gestures	26
Bluetooth On/Off	26
Image Capture	26

POWER

Power On	27
Power Off	28
Close-to-Off	28
Flip-to-Off	28

HOW TO WEAR

Check Your Fit	29
Try the Head Strap	30

CARE AND MAINTENANCE

Store the Frames	31
Maintain the Frames	32
Replacement Parts and Accessories	32
Return for Service and Repair	33

TROUBLESHOOTING

Try These Solutions First	34
Other Solutions	34

TECHNICAL INFORMATION

36

LABEL SYMBOLS

Symbols and Descriptions	38
--------------------------------	----

Important Safety Instructions:

1. Read these instructions.
2. Keep these instructions.
3. Heed all warnings.
4. Follow all instructions.
5. Do not use this device near bodies of water.
6. Clean only with a dry cloth.
7. Only use attachments/accessories specified by the manufacturer.
8. Unplug the device during lightning storms or when unused for long periods of time.
9. Refer all servicing to qualified personnel. Servicing may be required when the device has been damaged in any way, such as power-supply cord or plug is damaged, liquid has been spilled or objects have fallen onto the device, the device has been exposed to excessive rain or moisture, does not operate normally, or has been dropped.



WARNING/CAUTION

For Healthcare Professionals

Healthcare professionals must inform patients not to modify or discontinue their treatment plan without receiving instructions from the healthcare professional.

The responsible organization should identify, analyze, evaluate, and control any additional risks resulting from the device connected to IT networks, including other equipment.

Changes to the IT network could introduce new risks requiring additional analysis by the responsible organization. The changes include:

- changes in the IT network configuration
- connection of additional items to the IT network
- disconnecting items from the IT network
- update or upgrade of equipment connected to the IT network

NOTE: If a healthcare professional is unsure whether or not a patient may be eligible to use the device they should contact the manufacturer for inquiry at support@blinkframes.com

For Patients and Healthcare Professionals

Patients should not modify or discontinue their treatment plan without receiving guidance from a healthcare professional.

Ask your healthcare professional if contacts may be used when wearing the device.

Do not cover or obscure the cameras of the device during normal use and operation.

Although the device's own electromagnetic emissions are well below the levels permitted by the relevant standards, they may cause interference in other nearby devices, for example sensitive sensors.

If you need help using Blink Frames, contact your healthcare professional.

Do not use eye patches which have been previously used or have damage to their packaging.

- Use caution while driving and follow applicable laws regarding use of a mobile phone.
- Focus on your safety and that of others if you use the frames while engaging in any activity requiring your attention, e.g., while riding a bicycle or walking in or near traffic, a construction site or railroad, etc.
- Do NOT use the frames if they emit any loud unusual noise. If this happens, turn the frames off and contact Globe Biomedical, Inc. customer service at support@blinkframes.com
- Do NOT submerge or expose the frames for an extended period to water or wear them while participating in water sports, e.g., swimming, water skiing, surfing, etc.
- Remove frames immediately if you experience a warming sensation.



Contains small parts which may be a choking hazard. Not suitable for children under the age of 3.



This product contains magnetic material. Consult your physician on whether this might affect your implantable medical device.

- To reduce the risk of fire or electrical shock, do NOT expose this product to excessive rain, liquids or moisture.
- Do NOT expose this product to dripping or splashing, and do not place objects filled with liquids, such as vases, on or near the product.
- Keep the product away from fire and heat sources. Do NOT place naked flame sources, such as lighted candles, on or near the product.
- Do NOT make unauthorized alterations to this product.
- Use this product only with an agency-approved power supply that meets local regulatory requirements (e.g., UL, CSA, VDE, CCC).
- Do not expose products containing batteries to excessive heat (e.g. from storage in direct sunlight, fire, etc.).
- Do NOT wear the frames while charging.
- After each use, wipe both sides of the lenses and all parts of the frame with the cloth provided or a dry cloth.

The frames are intended to be worn by the person to whom they are prescribed only. Use on others, animals, or objects is forbidden. The manufacturer is not liable for any damage that may result from the improper use of the device.

It is not recommended to use the device in environments where temperatures are in excess of 110 °F or below 32 °F. Using the frames in temperatures in excess of 110 °F may damage the product, which could lead to harm of the user.

Do not charge the device in environments where the temperature is 110 °F/ 43 °C or higher.

When not in use, keep the Blink Frames on the provided charger or store in the provided bag to protect from direct sunlight.

Only lenses approved by Globe Biomedical, Inc. should be used with this product.

Not to be used for direct observation of the sun.

Not to be used for protection against artificial light sources.

Not to be used for use as eye protection against mechanical impact hazards.

Do not store or carry flammable liquids, gases, or explosive materials in the same compartment as your device, its parts, and/or accessories.

Do not charge or store your device in areas with potentially explosive atmospheres.

Keep the Product's charging device away from liquids. It is not water resistant and exposure to liquids could damage the charging device or other property and lead to injury.

Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the Blink Frames, including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.

May be unable to collect data when the device is in close vicinity to strong electromagnetic fields (e.g., electromagnetic anti-theft systems, metal detectors).

DO NOT wear your Blink Frames during a medical procedure (e.g., magnetic resonance imaging, diathermy, lithotripsy, cautery and external defibrillation procedures).

Contraindications

You should not use the Blink Frames based on the following:

- Blink Frames should not be prescribed to patients for whom the initial clinical setup fails to demonstrate valid eye imagery due to incompatibility between the sensing system of frames and the patient's facial geometry.

You may not be eligible for use of the Blink frames if:

- Your prescription lenses (sphere) are outside of the range of -5.00D to +5.00D in either eye.

The essential performance of Blink Frames is to capture images of the eye and the surrounding area, and to transfer the data into the Blink Frames software system.

If the essential performance of the frames is lost or degraded due to electromagnetic disturbances, the Blink Frames app will notify the user the next time the frames connect to Wi-Fi.

NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment to an outlet on a circuit that is different from that which the receiver is connected.
- Consult the manufacturer or an experienced radio/TV technician for help.

Changes or modifications not expressly approved by Globe Biomedical, Inc. could void the user's authority to operate this equipment.

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interferences received, including interference that may cause undesired operation.

This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.



This symbol means the product must not be discarded as household waste, and should be delivered to an appropriate collection facility for recycling. Proper disposal and recycling helps protect natural resources, human health, and the environment. For more information on disposal and recycling of this product, contact your local municipality, disposal service, or the shop where you bought this product.

The separate collection and recycling of your product or its battery, at the time of disposal, helps conserve natural resources and ensure that it is recycled in a manner that protects human health and the environment.

The sales package is cardboard and can be recycled.

Waste carton generally includes paper, carton, and cardboard packages.

Recycle according to local laws and regulations.

DO NOT attempt to remove the rechargeable lithium-ion battery from this product. Contact Globe Biomedical, Inc. for removal. **Please dispose of used batteries properly, following local regulations.** Do not incinerate.



Manufactured For: Globe Biomedical, Inc. 5225 Canyon Crest Dr Riverside CA



92507 Do not use if package is damaged.



This product does not contain natural rubber latex.



This product is provided non-sterile.



+45C

Security Information

! This product is capable of receiving security updates from Globe Biomedical, Inc. automatically when connected to Wi-Fi. In order to receive security updates, you must complete the product setup process in the Blink Frames app.

Apple, the Apple logo, iPad, iPhone, iPod, and Siri are trademarks of Apple Inc., registered in the U.S. and other countries. App Store is a service mark of Apple Inc.

Google and Google Play are trademarks of Google LLC.

Blink, the Blink Frames Logo  and the Blink Frames icon  are trademarks of Globe Biomedical, Inc.

Patented and Patent Pending technology, for more information visit www.blinkframes.com/patents

Globe Biomedical, Inc. Headquarters: 1-951-409-9695

©2023 Globe Biomedical, Inc. No part of this work may be reproduced, modified, distributed, or otherwise used without prior written permission.

The following license disclosures apply to certain third-party software packages used as components in the product.

1. The origin of this software must not be misrepresented; you must not claim that you wrote the original software. If you use this software in a product, an acknowledgment in the product documentation would be appreciated but is not required.
2. Altered source versions must be plainly marked as such, and must not be misrepresented as being the original software.
3. This notice may not be removed or altered from any source distribution.

Package: Mbed TLS 2.16.1

© ARM Limited. All rights reserved. The Mbed TLS package is subject to the license terms reprinted below.

Apache 2.0 License

Apache License

Version 2.0, January 2004

<http://www.apache.org/licenses/>

TERMS AND CONDITIONS FOR USE, REPRODUCTION, AND DISTRIBUTION

1. Definitions.

“License” shall mean the terms and conditions for use, reproduction, and distribution as defined by Sections 1 through 9 of this document.

“Licensor” shall mean the copyright owner or entity authorized by the copyright owner that is granting the License.

“Legal Entity” shall mean the union of the acting entity and all other entities that control, are controlled by, or are under common control with that entity. For the purposes of this definition, “control” means (i) the power, direct or indirect, to cause the direction or

management of such entity, whether by contract or otherwise, or (ii) ownership of fifty percent (50%) or more of the outstanding shares, or (iii) beneficial ownership of such entity.

“You” (or “Your”) shall mean an individual or Legal Entity exercising permissions granted by this License.

“Source” form shall mean the preferred form for making modifications, including but not limited to software source code, documentation source, and configuration files.

“Object” form shall mean any form resulting from mechanical transformation or translation of a Source form, including but not limited to compiled object code, generated documentation, and conversions to other media types.

“Work” shall mean the work of authorship, whether in Source or Object form, made available under the License, as indicated by a copyright notice that is included in or attached to the work (an example is provided in the Appendix below).

“Derivative Works” shall mean any work, whether in Source or Object form, that is based on (or derived from) the Work and for which the editorial revisions, annotations, elaborations, or other modifications represent, as a whole, an original work of authorship. For the purposes of this License, Derivative Works shall not include works that remain separable from, or merely link (or bind by name) to the interfaces of, the Work and Derivative Works thereof.

“Contribution” shall mean any work of authorship, including the original version of the Work and any modifications or additions to that Work or Derivative Works thereof, that is intentionally submitted to Licensor for inclusion in the Work by the copyright owner or by an individual or Legal Entity authorized to submit on behalf of the copyright owner. For the purposes of this definition, “submitted” means any form of electronic, verbal, or written communication sent to the Licensor or its representatives, including but not limited to communication on electronic mailing lists, source code control systems, and issue tracking systems that are managed by, or on behalf of, the Licensor for the purpose of discussing and improving the Work, but excluding communication that is conspicuously marked or otherwise designated in writing by the copyright owner as “Not a Contribution.”

“Contributor” shall mean Licensor and any individual or Legal Entity on behalf of whom a Contribution has been received by Licensor and subsequently incorporated within the Work.

2. Grant of Copyright License. Subject to the terms and conditions of this License, each Contributor hereby grants to You a perpetual, worldwide, non-exclusive, no-charge, royalty-free, irrevocable copyright license to reproduce, prepare Derivative Works of, publicly display, publicly perform, sub-license, and distribute the Work and such Derivative Works in Source or Object form.

3. Grant of Patent License. Subject to the terms and conditions of this License, each Contributor hereby grants to You a perpetual, worldwide, non-exclusive, no-charge, royalty-free, irrevocable (except as stated in this section) patent license to make, have made, use, offer to sell, sell, import, and otherwise transfer the Work, where such license applies only to those patent claims licensable by such Contributor that are necessarily infringed by their Contribution(s) alone or by combination of their Contribution(s) with the Work to which such Contribution(s) was submitted. If You institute patent litigation against any entity (including a cross-claim or counterclaim in a lawsuit) alleging that the Work or a Contribution incorporated within the Work constitutes direct or contributory patent infringement, then any patent licenses granted to You under this License for that Work shall terminate as of the date such litigation is filed.

4. Redistribution. You may reproduce and distribute copies of the Work or Derivative Works thereof in any medium, with or without modifications, and in Source or Object form, provided that You meet the following conditions:

You must give any other recipients of the Work or Derivative Works a copy of this License; and

You must cause any modified files to carry prominent notices stating that You changed the files; and

You must retain, in the Source form of any Derivative Works that You distribute, all copyright, patent, trademark, and attribution notices from the Source form of the Work, excluding those notices that do not pertain to any part of the Derivative Works; and

If the Work includes a “NOTICE” text file as part of its distribution, then any Derivative Works that You distribute must include a readable copy of the attribution notices contained within such NOTICE file, excluding those notices that do not pertain to any part of the Derivative Works, in at least one of the following places: within a NOTICE text file distributed as part of the Derivative Works; within the Source form or documentation, if provided along with the Derivative Works; or, within a display generated by the Derivative Works, if and wherever such third-party notices normally appear. The contents of the NOTICE file are for informational purposes only and do not modify the License. You may add Your own attribution notices within Derivative Works that You distribute, alongside or as an addendum to the NOTICE text from the Work, provided that such additional attribution notices cannot be construed as modifying the License.

You may add Your own copyright statement to Your modifications and may provide additional or different license terms and conditions for use, reproduction, or distribution of Your modifications, or for any such Derivative Works as a whole, provided Your use, reproduction, and distribution of the Work otherwise complies with the conditions stated in this License.

5. Submission of Contributions. Unless You explicitly state otherwise, any Contribution intentionally submitted for inclusion in the Work by You to the Licensor shall be under the terms and conditions of this License, without any additional terms or conditions. Notwithstanding the above, nothing herein shall supersede or modify the terms of any separate license agreement you may have executed with Licensor regarding such Contributions.

6. Trademarks. This License does not grant permission to use the trade names, trademarks, service marks, or product names of the Licensor, except as required for reasonable and customary use in describing the origin of the Work and reproducing the content of the NOTICE file.

7. **Disclaimer of Warranty.** Unless required by applicable law or agreed to in writing, Licensor provides the Work (and each Contributor provides its Contributions) on an "AS IS" BASIS, WITHOUT WARRANTIES OR CONDITIONS OF ANY KIND, either express or implied, including, without limitation, any warranties or conditions of TITLE, NON-INFRINGEMENT, MERCHANTABILITY, or FITNESS FOR A PARTICULAR PURPOSE. You are solely responsible for determining the appropriateness of using or redistributing the Work and assume any risks associated with Your exercise of permissions under this License.

8. **Limitation of Liability.** In no event and under no legal theory, whether in tort (including negligence), contract, or otherwise, unless required by applicable law (such as deliberate and grossly negligent acts) or agreed to in writing, shall any Contributor be liable to You for damages, including any direct, indirect, special, incidental, or consequential damages of any character arising as a result of this License or out of the use or inability to use the Work (including but not limited to damages for loss of goodwill, work stoppage, computer failure or malfunction, or any and all other commercial damages or losses), even if such Contributor has been advised of the possibility of such damages.

9. **Accepting Warranty or Additional Liability.** While redistributing the Work or Derivative Works thereof, You may choose to offer, and charge a fee for, acceptance of support, warranty, indemnity, or other liability obligations and/or rights consistent with this License. However, in accepting such obligations, You may act only on Your own behalf and on Your sole responsibility, not on behalf of any other Contributor, and only if You agree to indemnify, defend, and hold each Contributor harmless for any liability incurred by, or claims asserted against, such Contributor by reason of your accepting any such warranty or additional liability.

END OF TERMS AND CONDITIONS

Licensed under the Apache License, Version 2.0 (the "License"); you may not use this file except in compliance with the License.

You may obtain a copy of the License at [http://www.apache.org / licenses/LICENSE-2.0](http://www.apache.org/licenses/LICENSE-2.0)

Unless required by applicable law or agreed to in writing, software distributed under the License is distributed on an "AS IS" BASIS, WITHOUT WARRANTIES OR CONDITIONS OF ANY KIND, either express or implied.

See the License for the specific language governing permissions and limitations under the License.

The Blink Frames System Consists of the Following:

- Blink Frames
- Blink Frames Launch Pad
- Blink Frames Clinicians Portal, a browser-based software service where healthcare professional can view patient data.
- Blink Frames App, a smartphone application where the patient, healthcare professional, or guardian can review patient data.

With Blink Frames, clinicians and guardians who have received "shared" access will have the ability to review data related to an individual's ocular health. When the frames are worn by an individual, images of the eye and surrounding area are taken throughout the day and stored in the frames internal memory. Once the individual charges their frames, the data is securely offloaded to be reviewed.

Familiarize yourself thoroughly with the frames, software, and the operational procedures before use.

Information on Patch Wearing

Follow your healthcare professional's instructions on how to apply and if/when the patient should be wearing the provided patches. The eye patches are meant for one time use and should be discarded after being worn. When patching, use a new unused patch to cover the eye instructed by the healthcare professional. Use of other patches not produced/provided by Globe Biomedical, Inc. may negatively affect eye patch monitoring. For more information about the eye patches or how to order more patches please contact support@blinkframes.com

Support Materials

To learn to use the Blink Frames system, read this manual carefully. If you have problems using any part of the Blink Frames system, contact the organization from which you obtained the frames or Globe Biomedical, Inc. You can contact Globe Biomedical, Inc. at support@blinkframes.com

SWEAT AND WATER RESISTANCE

The frames are rated IPX3 water resistant. They are designed to be sweat and weather resistant against drops of water but are not meant to be submerged under water or exposed to excessive moisture.



NOTE: IPX3 is not a permanent condition, and resistance might decrease as a result of normal wear.

CAUTIONS:



- Do NOT use this device in a bathtub, shower or water-filled reservoir.
- Do NOT swim or participate in water sports with the frames.
- Do NOT submerge the frames in bodies of liquid.



LIMITED WARRANTY

The frames are covered by a limited warranty. Visit our website for details: <https://blinkframes.com/warranty-information>

To register your product, visit <https://blinkframes.com/registration> for instructions. Failure to register may affect your limited warranty rights.

LIFETIME

The expected service life of the frames is 2 years. The maintenance procedures described in this manual are required during the expected service life.

Inspect the frames for mechanical and functional damage annually. Contact the manufacturer if you detect any damage or deterioration.

MODEL/REFERENCE NUMBER LOCATIONS

The device product number, serial number, associated FCC ID, region of assembly, device input voltage, and IPX rating are located on the left temple.


The product name, model, and dimensional markings are located on the inner right temple.

FRAME HARDWARE

Charging pads are located on each side of the temple frames and are used for charging the frames when placed on the supplied charging device (Launch Pad).

The notification bars are located on each side of the frame and illuminate particular colors to inform the user of different frame operations/status.

Cameras are integrated into the frames to capture important data regarding the user's ocular health.



CONTENTS

Confirm that the following parts are included:



Blink Smart Frames



Launch Pad



Charging Cable



Head Strap



Cleaning Cloth



Storage Sleeve

NOTE: If any part of the product is damaged, do not use it.

For customer service, visit: <https://blinkframes.com/support>

The Blink Frames app lets you set up and control the frames from a smartphone running Android 10.0 or iOS 14 and above.

Using the app, you can setup and share data from a device to an authorized account, review data, review device details, and obtain new features.

DOWNLOAD THE BLINK FRAMES APP

1. On your mobile device, download the Blink Frames app.
 - A link to the "USER APP" may be located on the front of the Quick Start guide that came with your product packaging.



2. Follow the app instructions to setup your account.

ADDING YOUR FRAMES TO YOUR ACCOUNT

1. Make sure that your frames are charging on the provided Launch Pad see page 18.
2. In the Blink Frames app, tap the "add device" icon to add your frames.
3. Select your frames when they appear on the app to initiate frame setup.

NOTE: If the app cannot find the frames, see "Blink Frames app cannot find frames" on page 35.

CHARGE THE FRAMES

1. Connect the small end of the provided USB-C cable to the port on the back of the Launch Pad.



2. Connect the other end of the USB cable to an approved wall charger (not provided). **CAUTION:** Use this product only with an agency-approved, IEC 60601-1 power supply that meets local regulatory requirements (e.g., UL, CSA, VDE, CCC).
3. Fold the frames and set them on Launch Pad making sure the contact pads line up with the charge pins as shown below. When charging, the notification bar will change color. You may need to readjust the frames until the notification bar pulses white.

NOTE: During the initial charge/setup, the notification bar will pulse white. Once you finish setting up your frames, they will then pulse blue when placed on the charger. More information on notification bar colors may be found on page 23.



NOTE: Don't wear or attempt to wear the frames while charging. Remove any attached head strap prior to charging.


CHECK THE BATTERY LEVEL

- When connected over bluetooth, check the device page in the Blink Frames app.
- While charging, the notification bar will indicate charge level (see page 23).

CONNECTING TO YOUR FRAMES ON YOUR MOBILE DEVICE

1. On your mobile device enable bluetooth.

NOTE: The bluetooth feature is usually found in the Settings menu.



2. Open your Blink Frames app and navigate to the Summary page.
3. Place your frames on the Launch Pad charger.

NOTE: You can connect to your frames over bluetooth within the first 30 seconds of charging your frames to the Launch Pad. If you did not connect to the frames within this window simply remove the frames from the charger for 30-60 seconds and restart the process.



NOTE: During this time the notification bar will pulse blue.

4. In the Blink Frames app, tap the "Device" icon to add your frames.
5. Select your frames when they appear on the app.

NOTE: If this is your first time connecting to your frames, the app will walk you through the device setup process.

Once connected, the notification bar will glow solid blue.

These settings are only available when connected to your frame over bluetooth.

The below settings can be accessed by pressing the  icon on the device page

FRAME OPTIONS:

- **DISCONNECT:** Disconnect your mobile device from your frames.
- **AUTO-CONNECT:** Automatically connect to your frames when in the Device Menu.
- **DEVICE NAME:** Add a custom display name for your device.
- **WIFI:** Connect to an available Wifi network. From here, enter/select your Wifi SSID (network name) and then your network password under Wifi Password and select update.
- **DE-AUTHENTICATE FRAMES:** De-authenticating frames unlocks and removes the device from your account.
 - You will be asked to verify your sign-in credentials
 - All frame data will be deleted when the device is de-authenticated.
 - When placing the frames back on the charger the notification bar will revert to pulsing white.

NOTE: If you wish to add the device back onto your account, follow the instructions on page 19. Once the frames are authenticated to your account you may consider updating the device name.

- **WIPE FRAME MEMORY:** Will clear the captured and stored data on your frames.

NOTE: Clearing the internal frame memory permanently removes all stored image data collected on the device. This action can not be undone.

- **REBOOT FRAMES:** Will restart the hardware on the device.
 - For more information regarding why you may want to reboot your frames see page 34-35.

NOTE: Rebooting your frames will disconnect you from your device.

The notification bar reflects the *bluetooth*, battery charge, and error status of the frames.

Notification Bar



BLUETOOTH CONNECTION STATUS

Shows the *Bluetooth* connection status of mobile

LIGHT ACTIVITY	SYSTEM STATE
Slowly pulsing blue	Ready to connect
Quickly pulsing blue	Connecting
Solid blue	Connected

BATTERY STATUS

Shows the frames battery charge

LIGHT ACTIVITY	SYSTEM STATE
Slowly pulsing red	Battery is 0-20% while charging
Slowly pulsing orange	Battery is 21-80% while charging
Slowly pulsing green	Fully charged
Solid red	Uploading images 0-20% charged
Solid orange	Uploading images 21-80% charged
Solid green	Uploading images 81-100% charged

DEVICE STATUS

Shows the frame status.

LIGHT ACTIVITY	SYSTEM STATE
Slowly pulsing white	Frames are ready to be setup/authenticated.
Slowly pulsing yellow	Error - see page 35

IN-APP NOTIFICATIONS

The following is a list of in-app notification icons that the user may come across when interfacing with the blink frames app.



- Entered Wifi credentials were successful.



- Entered WiFi credentials were unsuccessful.



- Device has been updated.



- Blink Frames user sign-in credentials were successful.



- Blink Frames user sign-in credentials were unsuccessful.
- Blink Frames have successfully been de-authenticated
- The frames connected to are unlocked and not locked to your account.



- Connection via bluetooth action was successful.



- Connection via bluetooth was not successful.



- Device data has successfully uploaded.



- Offloaded data has been processed successfully and is ready for viewing in the app.

IN-APP NOTIFICATIONS

The following is a list of in-app notification icons that the user may come across when interfacing with the blink frames app.



- A new notification message is available.



- The frames are intentionally disconnecting.



- A self test has succeeded



- A self test has failed.



- There are no additional notifications for the frames to send.



- The frames detect an environment outside of their normal operating temperature.



- Alert that high screen time has been detected.




- Prolonged inactivity period(s) detected.



- Lower than average blink rate recorded.

TAP AND GESTURES



1. Tap and gesture controls can be changed in the app when connected to your frames over bluetooth.
2. By default, all tap and gesture settings are disabled.

CONTROL	FUNCTIONALITY
Bluetooth On/Off	Enables bluetooth connectivity for 10 seconds.
Image Capture	Enables/disables image capture


NOTES:

- When Image Capture is enabled the notification bar pulses green.
- When Image Capture is disabled, the notification bar pulses red.
- Image capture automatically defaults to Enabled post charging.

POWER ON

The frames will automatically power on after connecting to the Launch Pad and the notification bar will pulse white if not authenticated. If the frames have been authenticated to a user account they will pulse blue for up to 30 seconds.

When you remove the frames from the charger they will remain on until they are powered off, see page 28.



The notification bar is located on both temples.

NOTE: The frames ship with a partial charge and may need to charge before initial use.

POWER OFF

There are two ways to power off Blink Frames.


Close-to-Off



The notification bar flashes red.

Flip-to-Off

Flip the frames upside down for 2 seconds.



The notification bar flashes red.

NOTE:

- To power the frames back on, return them to an open, upright position. Notification bar will flash green to indicate normal operation.


CHECK YOUR FIT

Move your head up and down to check if the frames stay in place.

NOTE:

- If the frames don't feel comfortable and secure, see "Try the head strap" on page 30.

PROPER FIT



When the frames fit:

- The nose pads comfortably rest on the sides of your nose.
- The lenses comfortably rest over your eyes and don't move when you smile.

IMPROPER FIT



When the frames don't fit:

- The lenses might sit too low or too high on your face.
- The lenses might slide forward or too far from your face.

TRY THE HEAD STRAP

In the box you can find a head strap.

PLACING THE HEAD STRAP

1. Using the provided cleaning cloth, firmly push one end of the head strap along the back end of the temple frame. Repeat on the other side.



2. Check your fit (see page 29).

STORE THE FRAMES

The frames collapse for easy, convenient storage.

1. Fold the temples inward towards the lenses so the temples lay flat.
2. Place the frames in the provided pouch



NOTE:

- Make sure to charge the frames before using them after periods of prolonged storage.

MAINTAIN THE FRAMES

COMPONENT	CARE INSTRUCTIONS
Frames	<p>Prior to any maintenance, be sure to remove the frames. It is not recommended to wipe and clean any surfaces when the frames are being worn.</p> <p>After each use, wipe both sides of the lenses and all parts of the frame with the provided cleaning cloth or a dry cloth.</p> <p>In case of water exposure, dry the product thoroughly and clear the charging areas of residue or other debris.</p> <p>CAUTIONS:</p> <ul style="list-style-type: none"> Do NOT use any sprays near the frames. Do NOT use any solvents, chemicals, or cleaning solutions containing alcohol, ammonia, or abrasives. Do NOT allow liquids to spill into any openings.
Launch Pad	Wipe the Launch Pad with a dry cloth.
Headstrap	Wipe the headstrap with a dry cloth.

REPLACEMENT PARTS AND ACCESSORIES

If your frames or accessories are damaged or broken please contact support@blinkframes.com for information regarding service.

RETURN FOR SERVICE OR REPAIR

NOTE: Before contacting for service, write down the serial and model number of your frames and, if applicable, the service ID number displayed on the app.

Contact Globe Biomedical, Inc. for shipping instructions. Unless otherwise instructed by Globe Biomedical, Inc. there is no need to ship any accessories with the frames. Use a suitable cardboard or similar box with the appropriate packaging material to protect the frames during shipment. Return the frames using any shipping method that includes a proof of dispatch and delivery.

NOTE: For assistance in setting up, using, or maintaining the frames or to report unexpected operation or events, contact the manufacturer or manufacturer's representative.

RECYCLE



Do not dispose of the frames with household waste. Send it to an appropriate facility for recovery and recycling. The frames should be recycled as electronic waste.

The separate collection and recycling of your product or its battery at the time of disposal help conserve natural resources and ensure that it is recycled in a manner that protects human health and the environment.

The packaging can be recycled. Recycle according to local laws and regulations.

TRY THESE SOLUTIONS FIRST

If you experience problems with the frames:

- Connect the frames to power for 2 seconds and then disconnect (see page 20).
- Charge the battery (see page 20).
- Check the status lights (see page 23).
- Make sure your mobile device supports bluetooth connections.
- Make sure that your mobile device has the bluetooth connection set on.
- Download the Blink Frames app and run available software updates.
- Connect another mobile device (see page 21).

OTHER SOLUTIONS

If you could not resolve your issue, see the table below to identify symptoms and solutions to common problems. If you are unable to resolve your issue, contact Globe Biomedical, Inc. customer service.

Visit: blinkframes.com/support

PROBLEM	WHAT TO DO
Frames don't connect with mobile device	<p>On your device:</p> <ul style="list-style-type: none">• Turn the bluetooth feature off and then on. <p>Move your device closer to the frames and away from any interference or obstructions.</p> <p>Connect a different mobile device (see page 21).</p>
Blink Frames app doesn't work on device	<p>Make sure the Blink Frames app is compatible with your device.</p> <p>Uninstall the Blink Frames app on your device then reinstall the app (see page 19).</p>

PROBLEM	WHAT TO DO
Blink Frames app can't find frames	<p>Power on the frames (see page 27). If this is your first time connecting to the frames, verify that the notification bar is flashing white and follow the instruction on page 21.</p>
Intermittent Bluetooth connection	<p>In the Blink Frames app, run available software updates. Connect again (see page 21). Move the mobile device closer to the frames.</p>
Frames don't charge	<p>Ensure both ends of the charging cable are properly connected. If the frames have been exposed to high or low temperatures, let the frames return to room temperature and then try charging again. Try a different approved USB wall charger or wall socket. Try removing the frames from the Launch Pad and reconnecting them to the charging unit.</p>
Frames don't respond to touch control	<p>Make sure you are touching the correct touch surface (see page 26). Make sure your hands are dry. If wearing gloves, remove them before tapping the touch control surface. Make sure that tap is enabled via the Blink Frames App.</p>
Notification bar is pulsing yellow	<p>There may be a problem connecting to your wireless internet. Make sure your Wifi login credentials are entered in correctly. Make sure that your wireless internet is working properly.</p> <ul style="list-style-type: none"> • If there are problems, try unplugging your router, wait a minute, and plugging it back in, and then reconnect again. <p>If problems persist contact customer service for support by visiting blinkframes.com/support or emailing support@blinkframes.com.</p>

Operation environment:

Temperature: 0 °C to +45 °C (32 °F to 113°F)

Relative humidity: 30 % to 90 %

Atmospheric pressure: 800 hPa to 1,060 hPa

Storage environment:

Temperature: -10 °C to +55 °C (14 °F to 131 °F)

Relative humidity: 10 % to 95 %

Atmospheric pressure: 700 hPa to 1,060 hPa

Transport environment:

Temperature: -10 °C to +55 °C (14 °F to 131 °F)

Relative humidity: 10 % to 95 %

Atmospheric pressure: 500 hPa to 1,060 hPa

The device and its materials are compliant with RoHS Directive 2011/65/EU.

The device and its parts are not made of natural rubber latex.

Mode of operation: continuous

The manufacturer and serial number is located on the original product packaging. The model and frame dimensional markings are located on the inside face of the left temple frame.

There are no electrical connections from the frames to the patient. All the parts of the frames are applied parts and the frames have BF-type electric shock protection.

Power Supply

5V input power supply, 3.7V rechargeable battery, lithium-ion. Battery life, charge cycles vary by use and settings.

System Requirements for Blink Frames Clinician's Portal

Internet connection

Minimum web browser versions: Edge (v 88), Chrome (v 58), Firefox (v 53) and Safari (5.1.7)

System Requirements for Blink Frames App

Internet Connection

Android smart phone with USB OTG support, operating system v10 or newer or iPhone with operating system iOS 14 or newer.

IT Network Specifications

To transfer the data from the frames to a mobile device, the frames must be connected via bluetooth. The mobile device must be connected to the Internet. The frames can be used as a stand-alone without a bluetooth or Internet connection. The device is designed in such a way that network failures do not prevent the frames from working normally.

SYMBOLS AND DESCRIPTIONS

SYMBOL	DESCRIPTION
	Wi-Fi
	Choking hazard
	Catalogue number
	Customer support
	Date of manufacture
	Documentation
	Do not use if package is damaged
	Does not contain latex
	Downloads
	FCC certification
	Forum
	Green Dot trademark symbol
	Keep dry
	Magnetic interference
	Manufacturer
	Images
	Warnings/Cautions
	Waste of electrical and electronic equipment directive symbol
	Non-sterile
	Temperature Limit

SYMBOL	DESCRIPTION
 eIFU Indicator	Electronic instructions for use
 i	Refer to instructions for use
 !	Security information
 SN	Serial number
Rx Only (U.S.)	Federal law (U.S.) restricts this device to sale by or on the order of a physician or properly licensed practitioner.

